

Calibration Order Liquid Handling

PRO302F13v6en

Applicant	<i>Fill in</i>	PO nr/ Kp nr:	
Company		Invoicing	<i>Fill in if different</i>
Street		Att.	
Zip code		Street	
Place		Zip code	
Contact person		Place	
Departm./ kp		Delivery	<i>Fill in if different</i>
E-mail adres		Att.	
Phonenumber		Street	
Evt.mobiel		Zip code	
		Place	

Calibration method		Monochannel	Multichannel
<input type="checkbox"/> PIP Light	Calibration	2 volume / 4 measurements	2 volume / 4 measurements
<input type="checkbox"/> PIP Eco	Calibration & Prev. Maintenance	2 volume / 4 measurements	2 volume / 4 measurements
<input type="checkbox"/> PIP Classic	Calibration & Prev. Maintenance	2 volume /10 measurements	2 volume / 5 measurements
<input type="checkbox"/> PIP Pro	Calibration & Prev. Maintenance	3 volume /10 measurements	3 volume / 10 measurements
<input type="checkbox"/> PIP Care	Prev. Maintenance		
<small>Data below are included in our standard calibration method:</small>			
Specifications	ISO 17025 accr.	Recall	Language cert.
ISO 8655	Yes	6 months	English
			Recall sticker
			Calibration Date
			As Found
			No

Optional: **Only when listings differ from insurance plan (PRO302F14)**

Specifications	ISO 17025 accr.	Recall	Cert. Language	Recall Date	As Found.
<input type="checkbox"/> Producer* (add-on)	<input type="checkbox"/> No	<input type="checkbox"/> None	<input type="checkbox"/> Dutch		<input type="checkbox"/> Yes
<input type="checkbox"/> BIS		<input type="checkbox"/> 12 months	<input type="checkbox"/> German	<input type="checkbox"/> 12 months	
		<input type="checkbox"/>months	<input type="checkbox"/> French	<input type="checkbox"/>months	
			<input type="checkbox"/> Portuguese		
			<input type="checkbox"/> Spanish		

* Combination of manufacturer specifications with ISO 17025 accreditation is not always possible for small volumes

Decision rule:

We declare to agree that in case of an accredited calibration:

“If the sum of the absolute deviation and the limits of the measurement uncertainty are larger than the used specification limits (+/-) then the conclusion will be “Out of Specification” or “DOESN'T COMPLY”.”

Declaration of decontamination:

I hereby declare that all instruments in this calibration order are free of any radioactive, microbiological or Chemical contamination and that they can be processed by Avantor safely.

Date:.....

Authorized by:.....

Signature:.....



Calibration Order Liquid Handling

PRO302F13v6en

Instruments

Type	Volume	Count
Monochannel	Fix	
Monochannel	Variable	
Stepper		
8K-multichannel		
12K-multichannel		
Anders		